

VERIFICATION OF CALIBRATION REPORT

OCT 28 2010

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N <u>130126</u> ✓	
Name <u>Derrick Westlake</u>		ID# <u>5574</u> ✓	Date <u>10/25/10</u> ✓
A	Agency <u>Anchorage Airport Police</u>		Phone # <u>266-2411</u>
Instrument Location <u>AAPD - 6040 Dehavilland Ave., Anchorage, AK 99502</u>			
Alco S/N <u>X301583</u> ✓		Target Value <u>.078</u> ✓	High Pressure <u>1350 psi</u>
B	Alco Test Value Average <u>.074</u> ✓ 1 st Alco		<u>.073</u> ✓ 2 nd Alco
Signature <u>[Signature]</u>			
(OVER)			

BMB
12/9/10

(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:


- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

[Signature]
Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 22nd day of DEC, 2010.

[Signature] (Notary Seal Stamp)
Daniel L. Nicholas
Notary Public, State of Alaska
Commission Expires with Office

BT9 03/00



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130126

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130126

OCTOBER 25, 2010

OPERATOR'S NAME:

WESTLAKE/DERRICK/M

OPERATOR'S NUMBER: 5574

SUBJECT'S LAST NAME:

VERIFICATION OF CAL

SUBJECT'S FIRST NAME/MI :

L

O.L. #: L

DEPT/AGENCY: AHC6

CASE/REPORT:

TEST TYPE: U

ALCO TARGET VALUE: .078

ALCO S/N: X301583

— BREATH ANALYSIS —

.078 ADJUSTED FOR 29.46 in
ALCO TARGET .076 16:13
BLANK TEST .000 16:14
INTERNAL STANDARD VERIFIED 16:14
ALCO TV 29.46 in .074 16:15
BLANK TEST .000 16:16
SUBJECT SAMPLE .000 16:16
BLANK TEST .000 16:17
ALCO TV 29.46 in .073 16:18
BLANK TEST .000 16:19

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130126

OCTOBER 25, 2010
TIME 16:09

— DIAGNOSTIC CHECK —

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 02/20/01
HEATERS
SAMPLE CHAMBER: 49c
BREATH TUBE: 42c
BAROMETER: 29.46 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

! " # % & ' () * + , - . / 0 1 2 3 4 5 6 7 8 9 : ; < = > ? @ A B C D E F G
H I J K L M N O P Q R S T U V W X Y Z [\] ^ _ ` a b c d e f g h i j k l m n o
p q r s t u v w x y z { | } ~